



Processing UK NEQAS Full Blood Count survey samples through Sysmex analysers

UK NEQAS Full blood count survey samples do not contain material suitable for the counting of Platelets by optical or fluorescent methods on Sysmex haematology instruments.

Please process these samples in CBC or CBC+DIFF modes to ensure that you only report impedance platelet counts on Sysmex XE, XT and XN-series analysers